



Central Office Bids and Awards Committee  
**SUPPLEMENTAL BID BULLETIN NO. 01**

**SUPPLY, DELIVERY, INSTALLATION, AND COMMISSIONING OF VARIOUS LABORATORY EQUIPMENT AND ACCESSORIES**

Bidders are hereby notified of the clarifications, modifications, revisions and amendments to certain provisions of the Bid Documents particularly as published in the PhilGeps website. This shall form an integral part of the Bid Documents.

➤ **SECTION VII. TECHNICAL SPECIFICATIONS**

FROM	TO								
<p><b>LOT 1 - : One (1) Unit Potentiometric Titrator</b></p> <p><b>III. OTHER INCLUSIONS</b></p> <ol style="list-style-type: none"> <li>With minimum 5yrs existing and proof of installation of the same Titrator with Chloride ISE in the market.</li> <li>With trained technician or engineer from the manufacturer for the installation of the unit</li> <li>The proposed equipment can recognize and can import file formats from the existing unit.</li> <li>The supplier shall issue a certificate of undertaking that spare parts shall be available <u>10</u> years from the date of acceptance.</li> </ol>	<p><b>LOT 1 - : One (1) Unit Potentiometric Titrator</b></p> <p><b>III. OTHER INCLUSIONS</b></p> <ol style="list-style-type: none"> <li>With minimum 5yrs existing and proof of installation of the same Titrator with Chloride ISE in the market.</li> <li>With trained technician or engineer from the manufacturer for the installation of the unit</li> <li>The proposed equipment can recognize and can import file formats from the existing unit.</li> <li>The supplier shall issue a certificate of undertaking that spare parts shall be available 10 years from the date of acceptance.</li> <li><u>After sales service maintenance schedule should be indicated/stated.</u></li> </ol>								
<p><b>LOT 2 - One (1) Unit UV-Vis Spectrophotometer</b></p> <p><b>II. TECHNICAL SPECIFICATIONS</b></p> <table border="1"> <tr> <td><b>Monochromator</b></td><td>Blazed holographic grating with 1200 gr/mm, Czerny- Turner with 0.2 m focal length</td></tr> <tr> <td><b>Power Cord</b></td><td>UK Power Cord A 18AWG 2.5M 250V 10A</td></tr> </table> <p><b>IV. Inclusions</b></p> <ol style="list-style-type: none"> <li>The instrument installation, commissioning and initial operation.</li> <li>One (1) year warranty on equipment, parts and services against factory defects.</li> <li>After sales service maintenance schedule should be indicated/stated.</li> <li>The supplier shall issue a certificate of undertaking that spare parts shall be available 10 years from the date of acceptance.</li> </ol>	<b>Monochromator</b>	Blazed holographic grating with 1200 gr/mm, Czerny- Turner with 0.2 m focal length	<b>Power Cord</b>	UK Power Cord A 18AWG 2.5M 250V 10A	<p><b>LOT 2 - One (1) Unit UV-Vis Spectrophotometer</b></p> <p><b>II. TECHNICAL SPECIFICATIONS</b></p> <table border="1"> <tr> <td><b>Monochromator</b></td><td>Blazed holographic grating with 1200 gr/mm, Czerny- Turner with 0.2 m focal length <u>or equivalent or higher specifications</u></td></tr> <tr> <td><b>Power Cord</b></td><td>UK Power Cord A 18AWG 2.5M 250V 10A <u>or equivalent</u></td></tr> </table> <p><b>IV. Inclusions</b></p> <ol style="list-style-type: none"> <li>The instrument installation, commissioning and initial operation.</li> <li>One (1) year warranty on equipment, parts and services against factory defects.</li> <li>After sales service maintenance schedule should be indicated/stated.</li> <li>The supplier shall issue a certificate of undertaking that spare parts shall be available 10 years from the date of acceptance.</li> <li><u>With trained technician or engineer from the manufacturer for the installation of the unit</u></li> </ol>	<b>Monochromator</b>	Blazed holographic grating with 1200 gr/mm, Czerny- Turner with 0.2 m focal length <u>or equivalent or higher specifications</u>	<b>Power Cord</b>	UK Power Cord A 18AWG 2.5M 250V 10A <u>or equivalent</u>
<b>Monochromator</b>	Blazed holographic grating with 1200 gr/mm, Czerny- Turner with 0.2 m focal length								
<b>Power Cord</b>	UK Power Cord A 18AWG 2.5M 250V 10A								
<b>Monochromator</b>	Blazed holographic grating with 1200 gr/mm, Czerny- Turner with 0.2 m focal length <u>or equivalent or higher specifications</u>								
<b>Power Cord</b>	UK Power Cord A 18AWG 2.5M 250V 10A <u>or equivalent</u>								

➤ **SECTION VIII. CHECKLIST OF TECHNICAL AND FINANCIAL DOCUMENTS**

FROM	TO
(k) <u>FOR LOT 2 ONLY</u> : Certificate of after sales service maintenance schedule	(k) <u>FOR LOTS 1 &amp; 2 ONLY</u> : Certificate of after sales service maintenance schedule
(l) <u>FOR LOT 2 ONLY</u> : Certificate of undertaking that spare parts shall be available 10 years from the date of acceptance	(l) <u>FOR LOTS 1 &amp; 2 ONLY</u> : Certificate of undertaking that spare parts shall be available 10 years from the date of acceptance

  
**MARK ROY Q. MACANLALAY**  
Chairman, Bids and Awards Committee  
PCA Central Office